

## BRHNY-Series Bio-Burden Reduction Grade Nylon Plus+

BRHNY-Series Bio-Burden Reduction Grade Nylon Plus+ filter cartridges feature Nylon 6,6 membrane with an advanced positively-charged surface modification that is highly efficient in capturing submicronic particulate matter much finer than the stated mechanical rating. This offers a well-proven capability for highly efficient retention of haze, colloids, and color bodies. Specific to its use in medical applications, pyrogenic endotoxins are effectively removed as is well-documented in field use, industry journals, and laboratory data. Superior microbial retention is achieved to deliver a stable and consistent effluent. The BRHNY+ series offers a more cost-effective alternative to hollow-fiber cartridges in many high-purity applications. Cartridges are manufactured in a cleanroom environment and are flushed with 18 megaohm ultra-high purity water to ensure cleanliness, low extractables, and quick rinse-up for service use. Tolerant of repeated hot water sanitization and in-situ steam sterilization cycles for maximum service life. Each element is 100% integrity tested to Global Filter standards to assure consistent and optimal performance.

### Endotoxin Removal

Bacterial endotoxin is the pyrogen of greatest concern in the pharmaceutical and medical device industries. BRHNY+ filter elements have demonstrated capability to remove bacterial endotoxin to below a 0.005 EU/milliliter detection limit at all data points in independent testing.

### Microbial Retention Performance

Rating	Challenge Microbe	Log Reduction Value (LVR)
0.05µ	<i>Brevundimonas diminuta</i>	>10.1
0.10µ		>9.1
0.2µ		>9.0
0.45µ	<i>Serratia marcescens</i>	>9.0

\*Independently tested in accordance with ASTM F838

### Typical Applications


- Medical device reprocessing
- Water for Injection (WFI)
- Endotoxin removal
- Pyrogen removal

### Ordering Information

BRHNY+	Rating(µ)	A	Length	C	End Cap Style	O-Rings/Gaskets	-	Adders
	0.05		10" (25.4cm)		3 = 222 w/Fin	E = EPDM		CS = 316SS Compression Spring
	0.10		20" (50.8 cm)		4 = 222 w/Flat Cap	S = Silicone		I = Stainless Steel Insert
	0.2		30" (76.2 cm)		6 = 226 w/Flat Cap	V = Viton®		
	0.45		40" (101.6 cm)		7 = 226 w/Fin			
					16 = 213 Internal O-Ring			
					28 = 222 3-Tabs w/ Fin			



### Flow Rate vs Pressure Drop



### Construction Materials

**Membrane**.....Positively-charged Nylon 6,6 on polyester substrate  
**Support Media**.....Polypropylene  
**End Caps**.....Polypropylene  
**Center Core**.....Polypropylene  
**Outer Support Cages**.....Polypropylene  
**O-Rings/Gaskets**.....EPDM, Silicone, Viton®

### Dimension (Nominal)

**Length** .....10 to 40 in (25.4 to 101.6 cm)  
**Outside Diameter** .....2.7 in (7.06 cm)

### Operating Conditions

**Change Out ΔP (recommended)** ..... 35 PSID (2.4 bar)  
**Temperature (max)** ..... 176°F (80°C)  
**Differential Pressure (max)** .....50 PSID at 68°F (3.4 bar at 20°C)

### Sanitization/Sterilization

**Filtered Hot Water** ..... 176°F (80°C) for 30 min  
**Steaming Sterilization**..... 250°F (121°C) for 30 min multiple cycles

**Chemicals:** Cartridges are compatible with most chemical sanitizing agents.

**Note:** Stainless steel insert option required for all cartridges being hot water sanitized or steam sterilized.

### Toxicity

All polypropylene components meet the specifications for biological safety per USP Class VI – 121°C for plastics.

### Food Safety Compliance

Materials of construction comply with FDA regulations for food and beverage contact use as detailed in the US Code of Federal Regulations, 21CFR. Materials used to produce filter media and hardware are deemed safe for use in contact with foodstuffs in accordance with EU Directives 1935/2004, and/or 10/2011.

### AAMI Standard ST108 Compliance

The BRHNY+ is a key component of water systems required to be compliant to AAMI Standard ST108: Water for the Reprocessing of Medical Devices. The BRHNY+ delivers highly efficient scavenging of microbes and endotoxin when used in a properly-designed system maintained to ST108 recommended practices. This assures continued compliance of the system to meet the highest standards of cleanliness and user confidence.